

# A Synchrotron-Based Fourier-Synthesis Custom-Coherence Illuminator

#### **Patrick Naulleau**

Kenneth Goldberg, Phil Batson, Paul Denham, and Senajith Rekawa

Center for X-Ray Optics, Lawrence Berkeley National Laboratory



In collaboration with the Virtual National Laboratory



**Supported by the EUV LLC** 

#### **Outline**



- Overview of spatial coherence
- Issues with coherence and lithography at the ALS
- Fourier-synthesis illuminator
- Lithographic demonstration of coherence control






#### What makes a source incoherent?



- In real space, a low-coherence source can be thought of as a source with many independent point radiators
- The frequency-space equivalent description is a source with many independent plane-wave radiators




- "independent" means unable mutually incoherent (add in field).
  - Thus illumination divergence is a necessary but not sufficient condition for reduced coherence
  - "independence" can be guaranteed be ensuring that the "source elements" do not coexist in time





# Optimal imaging performance generally requires illumination partial coherence (σ>0.5)



$$\sigma = \theta_{c} / \theta_{p}$$

Specific case of critical illuminator with incoherent source

$$\sigma = Res / L_c$$

Generally,  $\sigma$  is ratio of imaging optic resolution to illumination coherence width





### Coherent illumination extremely susceptible to phase errors



aerial images Calculated ideal resist Images in **1-μm** 1-μm In focus In focus defocus defocus **Partially Coherent** Coherent  $(\sigma = 0.7)$ 




100-nm contacts imaged with 0.1-NA system at  $\lambda$  = 13.4nm




#### What is EUV lithography?





### Synchrotron-based advanced lithography tool at the ALS






- Static imaging in 100 μm subfield at wafer.
- Cover full static arc-field subfield-by-subfield.
- Use standard EUV reflection masks.
- 0.1-NA pre-production, diffraction-limited optic
- operating wavelength= 13.4 nm



#### Relevant lithography requires undulatorbeam coherence to be reduced








Measured coherence
7 µm


- Object Res = 0.4 μm
- $\sigma \approx 0.4/7 = 0.057$
- intrinsic coherence~10x too high

Two-pinhole interference with 5 μm separation\*












#### **Reducing coherence**




#### **Scatter plate**



- Method often used at visible wavelengths
- Difficult to implement and low efficiency in EUV and Soft-X-Ray regime

#### **Scanning illuminator**




mutually incoherent plane waves

- Readily implemented at short wavelengths
- High efficiency
- Facilitates programmable coherence capability



### Programmable scanning illuminator implemented in ALS lithography tool






instantaneous spatial spectrum


time-integrated spatial spectrum








In vacuum 2D angle scanner



Collector assembly




Scanning illuminator enables in situ

control of illumination partial coherence

Faurck naureau 10

### Illuminator-generated EUV pupil fills recorded using in situ CCD-based pupil-fill monitor






- Critical illumination
- Pupil fill is visualization of source spatial spectrum
- Source spectrum and spatial coherence form Fouriertransform pair




## Interference effects limit the usefulness of coherent illumination for lithography








## Partially-coherent illumination ( $\sigma$ = 0.7) dramatically improves performance






















# Direct comparison of coherence vs partial coherence near coherent resolution limit





Patrick Naulleau









#### **Modeling results consistent with printing**






### 50-nm lines and spaces well beyond conventional Rayleigh limit





50-nm lines and spaces

Pupil image for 50-nm lines under conventional  $\sigma = 0.7$  illumination






Rayleigh limit =  $(0.61\lambda)/NA$ = 82 nm




#### Dipole illumination enables Rayleigh limit to be surpassed





50-nm lines and spaces

Pupil image for 50-nm lines under dipole illumination










# Dipole illumination enhances resolution in only one direction
















# Annular illumination enhances resolution isotropically at the cost of effectiveness















# Programmable illuminator enables modeling of esoteric systems





6-Stripe fill


Disk fill

















70 nm



# Next-generation EUV lithography station requires more advanced illuminator design

- 0.3 NA optic with sub-30-nm resolution
- 600x200-um field of view






- 3x increase in NA requires 3x increase in scan range to maintain same σ
- 6x increase in field of view required, while maintaining stationarity and uniformity
- 10x increase in speed desired for improved stability



### Multi-element cylindrical scanner meets design requirements





#### **Summary**



- Coherence control crucial to advanced imaging applications
- Scanning illuminators well-adapted to synchrotron use
- Effectiveness of scanning illuminator has been demonstrated through lithographic studies
  - 50-nm lines-space printing demonstrated with 0.1-NA optic
  - Programmable capabilities allow illumination to be tuned to pattern
- Scanning illuminator design extended to 0.3-NA optic





#### **Acknowledgements**



#### **LBNL**

Left to right: Patrick Naulleau, Dave Richardson, Rene Delano, Senajith Rekawa, Keith Jackson, Jeffrey Bokor, Ron Tackaberry, Kenneth Goldberg, Farhad Salmassi, Paul Denham, Brian Hoef, Drew Kemp, Phillip Batson, Gideon Jones

Not pictured: Erik Anderson, Bruce Harteneck, Deirdre Olynick



#### LLNL: POB mods/instrumentation, mask fabrication

Sherry Baker, Ken Blaedel, Henry Chapman, Layton Hale, Paul Mirkarimi, Nhan Nguyen, Eberhard Spiller, Chris Walton

#### **SNL: Resist and SEM support**

Donna O'Connell, Kevin McDonald

**Intel: Mask fabrication** 

Pei-Yang Yan, Guojing Zhang



